



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 26, 2012 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 2 ADDRESS: I-5S MP141

LOCATION: SeaTac SB 27 E2

COUNTY: King

MAKE: Toledo

TYPE: Trk Scale

ACCURACY CLASS: IIII

MODEL: 8530

ID. NO: 4337886-4NV

CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: Digital

PLATFORM 12x20

East2

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.


Signed by:

David Cromer

Date:

Aug 20, 2012

Print name: David Cromer



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on Aug. 20, 2012)

Patricia A. Geppert
(Signature of notary public)

Printed Name of Notary

My appointment expires Sept. 4, 2015